

(Garuda)

Form NSC/TISI 2

Certificate No. 12132/C0217

Certificate of Accreditation Laboratory

By virtue of National Standardization Act B.E. 2551 (2008)

Secretary-General, Thai Industrial Standards Institute

Issue this Certificate for

Inctech Metrological Center co.,ltd

Laboratory address :

11/22, Soi Saimai 56/1, SaiMai Rd., SaiMai, SaiMai,Bangkok

This laboratory is accredited for calibration in accordance with
the Thai Industrial Standard TIS 17025-2548 (2005) (ISO/IEC 17025:2005)
General Requirements for the Competence of Testing and Calibration Laboratories.

Accreditation No. CALIBRATION 0217

The scope of accreditation is as annexed hereto.

Issue Date : 23 August B.E. 2555 (2012)

Valid until : 22 August B.E. 2558 (2015)

Signature :

Translation approved



(Yannapat Uthongsap)

Director,

Office of the National Accreditation Council

Date: 6 October 2014

(Urit Srinongkote)

Secretary - General

Thai Industrial Standards Institute

Date of Initial Issue 23 August B.E. 2555 (2012)

Ministry of Industry, Thai Industrial Standards Institute



Translation Note: In the event of doubt or misunderstanding, the original in Thai shall be the authoritative.

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Laboratory Name : Inctech Metrological Center co.,ltd

Address : 11/22, Soi Saimai 56/1, SaiMai Rd., SaiMai, SaiMai, Bangkok

Accreditation No. : Calibration 0217

Laboratory Status ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Rage/Item	Calibration and Measurement Capability*	Method/Technique
1. Electrical	Measuring instrument DC voltage 0 mV to < 320 mV 320 mV to < 3.2 V 3.2 mV to < 32 V 32 V to < 320 V 32 V to < 1 kV AC voltage 10 mV to < 32 mV @ 40 Hz to < 3 kHz @ 3 kHz to < 10 kHz @ 10 kHz to < 30 kHz @ 30 kHz to < 50 kHz @ 50 kHz to 100 kHz 32 mV to < 320 mV @ 40 Hz to < 3 kHz @ 3 kHz to < 10 kHz @ 10 kHz to < 30 kHz @ 30 kHz to < 50 kHz @ 50 kHz to 100 kHz	70 μ V/V + 7.7 μ V 70 μ V/V + 50 μ V 77 μ V/V + 0.50 mV 77 μ V/V + 5.3 mV 71 μ V/V + 24 mV 0.53 mV/V + 0.12 mV 0.52 mV/V + 0.15 mV 0.84 mV/V + 0.28 mV 1.4 mV/V + 0.56 mV 2.5 mV/V + 1.5 mV 0.51 mV/V + 27 μ V 0.50 mV/V + 33 μ V 0.76 mV/V + 61 μ V 1.2 mV/V + 0.17 mV 2.4 mV/V + 0.30 mV	In - house method : CP - DMM – 01A by direct measurement with multi – function calibrator In - house method : CP - DMM – 01A by direct measurement with multi – function calibrator
* express as an uncertainty (\pm), providing a level of confidence of approximately 95%			

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Range/Item	Calibration and Measurement Capability*	Method/Technique
1. Electrical (cont.)	Measuring Instrument AC voltage (cont.) 320 mV to < 3.2 V @ 40 Hz to < 3 kHz @ 3 kHz to < 10 kHz @ 10 kHz to < 30 kHz @ 30 kHz to < 50 kHz @ 50 kHz to 100 kHz 3.2 V to < 32 V @ 40 Hz to < 3 kHz @ 3 kHz to < 10 kHz @ 10 kHz to < 30 kHz @ 30 kHz to < 50 kHz @ 50 kHz to 100 kHz 32 V to < 105 V @ 40 Hz to < 3 kHz @ 3 kHz to < 10 kHz @ 10 kHz to < 30 kHz @ 30 kHz to < 50 kHz @ 50 kHz to 100 kHz 105 V to < 320 V @ 55 Hz to < 100 Hz @ 100 kHz to < 1 kHz @ 1 kHz to < 3 kHz @ 3 kHz to < 10 kHz @ 10 kHz to < 20 kHz @ 20 kHz to 30 kHz	0.51 mV/V + 0.26 mV 0.50 mV/V + 0.33 mV 0.76 mV/V + 0.58 mV 1.2 mV/V + 1.7 mV 2.4 mV/V + 3.2 mV 0.51 mV/V + 2.7 mV 0.72 mV/V + 3.3 mV 0.97 mV/V + 6.1 mV 1.9 mV/V + 17 mV 4.1 mV/V + 39 mV 0.51 mV/V + 9.4 mV 0.72 mV/V + 12 mV 0.97 mV/V + 20 mV 1.9 mV/V + 39 mV 4.1 mV/V + 0.13 V 0.61 mV/V + 28 mV 0.61 mV/V + 28 mV 0.95 mV/V + 28 mV 0.95 mV/V + 40 mV 1.4 mV/V + 58 mV 1.8 mV/V + 75 mV	In - house method : CP - DMM – 01A by direct measurement with multi – function calibrator
* express as an uncertainty (\pm), providing a level of confidence of approximately 95%			

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Range/Item	Calibration and Measurement Capability*	Method/Technique
1. Electrical (cont.)	<p>Measuring Instrument</p> <p>AC voltage (cont.)</p> <p>320 V to < 800 V</p> <p>@ 40 Hz to < 100 Hz</p> <p>@ 100 kHz to < 1 kHz</p> <p>@ 1 kHz to < 3 kHz</p> <p>@ 3 kHz to < 10 kHz</p> <p>@ 10 kHz to 20 kHz</p> <p>800 V to 1 050 V</p> <p>@ 40 Hz to < 100 Hz</p> <p>@ 100 Hz to < 1 kHz</p> <p>@ 1 kHz to < 3 kHz</p> <p>AC current</p> <p>32 μA to < 32 μA</p> <p>@ 55 Hz to < 3 kHz</p> <p>@ 3 kHz to < 5 kHz</p> <p>32 μA to < 3.2 mA</p> <p>@ 55 Hz to < 3 kHz</p> <p>@ 3 kHz to < 5 kHz</p> <p>3.2 mA to < 32 mA</p> <p>@ 55 Hz to < 3 kHz</p> <p>@ 3 kHz to < 5 kHz</p> <p>32 mA to < 320 mA</p> <p>@ 55 Hz to < 3 kHz</p> <p>@ 3 kHz to < 5 kHz</p> <p>320 mA to < 3.2 A</p> <p>@ 55 Hz to < 3 kHz</p> <p>3.2 A to < 10 A</p> <p>@ 55 Hz to < 3 kHz</p>	<p>0.61 mV/V + 75 mV</p> <p>0.61 mV/V + 75 mV</p> <p>0.95 mV/V + 75 mV</p> <p>0.95 mV/V + 0.13 V</p> <p>1.5 mV/V + 0.19 V</p> <p>0.61 mV/V + 0.16 V</p> <p>0.61 mV/V + 0.16 V</p> <p>0.95 mV/V + 0.16 V</p> <p>0.89 mA/A + 1.1 μA</p> <p>1.3 mA/A + 2.1 μA</p> <p>0.89 mA/A + 3.5 μA</p> <p>1.3 mA/A + 7.0 μA</p> <p>0.89 mA/A + 13 μA</p> <p>1.3 mA/A + 15 μA</p> <p>1.2 mA/A + 0.24 mA</p> <p>2.7 mA/A + 0.47 Ma</p> <p>1.2 mA/A + 0.63 mA</p> <p>2.4 mA/A + 3.7 mA</p>	<p>In - house method : CP - DMM – 01A</p> <p>by direct measurement with multi – function calibrator</p> <p>In - house method : CP - DMM – 01A by direct measurement with multi – function calibrator</p>
* express as an uncertainty (\pm), providing a level of confidence of approximately 95%			

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Range/Item	Calibration and Measurement Capability*	Method/Technique
1. Electrical (cont.)	<p>Measuring instrument</p> <p>DC current</p> <p>0 μA to < 320 μA</p> <p>320 μA to < 3.2 mA</p> <p>3.2 mA to < 32 mA</p> <p>32 mA to < 320 mA</p> <p>0.320 A to < 3.2 A</p> <p>3.2 A to 10 A</p> <p>Resistance 4 wire</p> <p>10 Ω to < 40 Ω</p> <p>40 Ω to < 400 Ω</p> <p>400 Ω to < 4 kΩ</p> <p>4 kΩ to 40 kΩ</p> <p>Resistance 2 wire</p> <p>40 kΩ to < 400 kΩ</p> <p>400 kΩ to < 4 MΩ</p> <p>4 MΩ to < 40 MΩ</p> <p>40 MΩ to 100 MΩ</p> <p>Quart stop watch</p> <p>Quart time base Oscillator</p> <p>32 768 (=2¹⁵) Hz</p>	<p>2.6 mA/A + 20 nA</p> <p>2.6 mA/A + 0.13 μA</p> <p>2.6 mA/A + 1.7 μA</p> <p>3.3 mA/A + 18 μA</p> <p>12.1 mA/A + 1.9 mA</p> <p>11.8 mA/A + 1.4 mA</p> <p>0.29 mΩ/Ω + 12 mΩ</p> <p>0.24 mΩ/Ω + 24 mΩ</p> <p>0.18 mΩ/Ω + 93 mΩ</p> <p>0.29 mΩ/Ω + 1.1 Ω</p> <p>0.29 mΩ/Ω + 11 Ω</p> <p>0.75 mΩ/Ω + 0.3 kΩ</p> <p>1.8 mΩ/Ω + 2.4 kΩ</p> <p>3.1 mΩ/Ω + 64 kΩ</p> <p>0.76 ms/s</p>	<p>In - house method : CP - DMM – 01A by direct measurement with multi – function calibrator</p> <p>In - house method : CP - DMM – 01A by direct measurement with multi –function calibrator</p> <p>In – house method : CP-ELE-01A by direct measurement with frequency counter</p>
* express as an uncertainty (\pm), providing a level of confidence of approximately 95%			

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Range/Item	Calibration and Measurement Capability*	Method/Technique
1. Electrical (cont.)	Electronic counter 1 to 9 999 count	0.29 count	In-house method : CP-ELE-03A by direct measurement with multi-function calibrator
	Digital tachometer Photo Type 6 rpm to 999.9 rpm 1 000 rpm to 99 999 rpm	0.060 rpm 0.58 rpm	In-house method : CP-ELE-02A by direct measurement with multi-function calibrator
	Digital tachometer Contact type 6 rpm to 999.9 rpm 1 000 rpm to 99 999 rpm	0.060 rpm 0.58 rpm	In-house method : CP-ELE-02A by direct measurement with multi-function calibrator (*display unit only , not include effect of sensor)
	Temperature indicator Thermocouple Type J -200 °C to 700 °C > 700 °C to 1200 °C	0.40 °C 0.39 °C	In - house method : CP-TEM-01A by direct measurement with documenting process
	Type T 0 °C to 400 °C	0.33 °C	calibrator based on EA10/11
	Type K -200 °C to 300 °C > 300 °C to 700 °C > 700 °C to 1 370 °C	0.40 °C 0.37 °C 0.38 °C	
	Type R 0 °C to 900 °C > 900 °C to 1 750 °C	1.2 °C 0.70 °C	
* express as an uncertainty (\pm), providing a level of confidence of approximately 95%			

Certificate No. 12132/C0217

Laboratory Status ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Ans

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Rage/Item	Calibration and Measurement Capability*	Method/Technique
2.Dimension (cont.)	Dial gauge (Digital and Analog) 0 mm to 10 mm >10 mm to 30 mm >30 mm to 50 mm Dial test indicator (Digital and Analog) 0 mm to 1 mm Bore gauge (Digital and Analog) 0 mm to 1.4 mm. Height gauge (Digital and Analog) 0 mm to 150 mm >150 mm to 300 mm >300 mm to 450 mm >450 mm to 600 mm Plain plug gauge 0.1 mm to 15 mm >15 mm to 100 mm >100 mm to 200 mm (0.1 mm to 200 mm) Plain Ring Gauge 1 mm to 15 mm >15 mm to 90 mm >90 mm to 250 mm (1 mm to 250 mm)	6.4 µm 6.5 µm 6.6 µm 6.0 µm 5.9 µm 8 µm 11 µm 14 µm 18 µm 1.1 µm 2.1 µm 2.2 µm 1.2 µm 1.7 µm 1.8 µm	In - house method : CP-DIA-01A based on JIS B 7503 : 1997 In - house method : CP-DIA-02A based on JIS B 7533 : 1990 In - house method : CP-DIA-04A based on JIS B 7515 : 1982 In - house method : CP-DIA-05A based on JIS B 7517 : 1993 In - house method : CP-DIA-06A based on ISO 286-1(E):1988 direct measurement by ULM using knife edged anvils probe In - house method : CP-DIA-07A And CP-DIA-08A based on ISO 286-1(E):1988 direct measurement by ULM using ruby ball probe and L-shaped probe
* express as an uncertainty (±), providing a level of confidence of approximately 95%			

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Rage/Item	Calibration and Measurement Capability*	Method/Technique
2.Dimension (cont.)	Pin Gauge /Thread Measuring wire		In - house method : CP-DIA-09A based on ISO 286-1(E):1988
	0.1 mm to 1 mm	1.1 μ m	direct measurement by
	>1 mm to 5 mm	1.1 μ m	ULM using knife edged
	>5 mm to 10 mm	1.2 μ m	anvils probe
	>10 mm to 20 mm	1.2 μ m	
	>20 mm to 30 mm	1.2 μ m	
	>30 mm to 40 mm	1.2 μ m	
	>40 mm to 50 mm	1.2 μ m	
	(0.1 mm to 50 mm)		
	Thread plug gauge		In - house method : CP-DIA-10A based on EA-10/10 : 1999
	M 1	1.9 μ m	direct measurement by
	\leq M 3	1.9 μ m	ULM using set 3-wires on holders
	\leq M 30	1.9 μ m	
	\leq M 68	2.1 μ m	
	(1 mm to 68 mm)		
	Thread ring gauge		In - house method : CP-DIA-11A based on EA-10/10 : 1999
	M 3	2.1 μ m	direct measurement by
	\leq M 30	2.1 μ m	ULM using T-shape probe
	\leq M 68	2.2 μ m	
	(3 mm to 68 mm)		
* express as an uncertainty (\pm), providing a level of confidence of approximately 95%			

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Range/Item	Calibration and Measurement Capability*	Method/Technique
3. Mass	Electronic balance and Mechanical balance up to 20 g > 20 g to 40 g > 40 g to 60 g > 60 g to 100 g > 100 g to 200 g > 200 g to 300 g > 300 g to 400 g > 400 g to 500 g > 500 g to 600 g > 600 g to 700 g > 700 g to 800 g > 800 g to 900 g > 900 g to 1 kg > 1 kg to 10 kg > 10 kg to 20 kg > 20 kg to 400 kg > 400 kg to 600 kg > 600 kg to 800 kg > 800 kg to 900 kg > 900 kg to 1 000 kg	0.13 mg 0.25 mg 0.26 mg 0.38 mg 0.65 mg 1.5 mg 1.6 mg 1.9 mg 2.1 mg 2.3 mg 2.5 mg 2.7 mg 3.0 mg 0.13 g 0.53 g 82 g 83 g 84 g 85 g 86 g	In - house method : CP-BAL-01A based on UKAS LAB 14 : 2006
* express as an uncertainty (\pm), providing a level of confidence of approximately 95%			

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Range/Item	Calibration and Measurement Capability*	Method/Technique
4. Mechanical	Pressure measuring instrument (Analog and digital) Pneumatic type 0 kPa to 200 kPa	0.19 kPa	In – house method : CP-PRE-01A based on DKD-R6-1,2003
	Water type 0 MPa to 70 MPa	58 kPa	In – house method : CP-PRE-02A based on DKD-R6-1,2003
	Vacuum measuring instrument (Analog and digital) -95 kPa to 0 kPa	0.33 kPa	In – house method : CP-PRE-03A based on DKD-R6-1,2003
	Hand torque tools : Screw Driver - Type I Class D , E - Type II Class D , E, F 0.5 N·M to 20 N·M	2.0 %	In – house method : CP-TOR-01A based on ISO 6789-2003
	Hand torque tools : Torque wrench - Type I Class A, B, C - Type II Class A, B, C, G 10 N·M to 1500 N·M	2.0 %	In – house method : CP-TOR-01A based on ISO 6789-2003
* express as an uncertainty (\pm), providing a level of confidence of approximately 95%			

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Rage/Item	Calibration and Measurement Capability*	Method/Technique
5. Temperature	Temperature sensor		In - house method : CP-TEM-03A by comparison with thermometer standard
	Thermocouple		
	Type J		
	-20 °C to 100 °C	0.55 °C	
	>100 °C to 200 °C	0.89 °C	
	>200 °C to 400 °C	1.9 °C	
	>400 °C to 600 °C	2.9 °C	
	Type K		
	-20 °C to 100 °C	0.73 °C	
	>100 °C to 200 °C	0.99 °C	
	>200 °C to 300 °C	1.4 °C	
	>300 °C to 400 °C	1.7 °C	
	>400 °C to 500 °C	2.1 °C	
	>400 °C to 600 °C	2.4 °C	
	Type T		In - house method : CP-TEM-30A by comparison with thermometer standard
	-20 °C to 150 °C	1.2 °C	
	>150 °C to 250 °C	2.3 °C	
	Resistance temperature detector		
	(Pt 100 Ω) 2, 3, 4 wire		
	-20 °C to 200 °C	0.25 °C	
	>200 °C to 300 °C	1.0 °C	
	>300 °C to 400 °C	1.3 °C	
	>400 °C to 500 °C	1.6 °C	
	>500 °C to 600 °C	1.9 °C	

* express as an uncertainty (\pm), providing a level of confidence of approximately 95%

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Range/Item	Calibration and Measurement Capability*	Method/Technique
5. Temperature (cont.)	Temperature indicator with sensor Thermocouple Type E, J, K, N, T -20 °C to 100 °C >100 °C to 200 °C >200 °C to 300 °C >300 °C to 400 °C Thermocouple Type E, J, K, N >400 °C to 500 °C >500 °C to 600 °C Resistance temperature detector (Pt 100 Ω) 2, 3, 4 wire -20 °C to 200 °C >200 °C to 300 °C >300 °C to 400 °C >400 °C to 500 °C >500 °C to 600 °C Dial Thermometer -20 °C to 200 °C >200 °C to 600 °C	0.44 °C 0.76 °C 1.4 °C 1.9 °C 2.4 °C 2.9 °C 0.13 °C 0.38 °C 0.57 °C 0.86 °C 0.89 °C 0.60 °C 1.1 °C	In - house Method : CP-TEM-13A by comparison with thermometer standard In - house method : CP-TEM-17A by comparison with thermometer standard In - house method : CP-TEM-05A by comparison with thermometer standard
* express as an uncertainty (±), providing a level of confidence of approximately 95%			

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Rage/Item	Calibration and Measurement Capability*	Method/Technique
5. Temperature (cont.)	Liquid in glass thermometer		In - house method : CP-TEM-10A by comparison with thermometer standard
	Total immersion		
	-20 °C to 100 °C	0.60 °C	
	Partial immersion		
	-20 °C to 100 °C	0.61 °C	
	Digital thermo – hygrometer		In – house method : CP-TEM-09A based on NPL : A guide to the measurement of humidity Temperature calculated @ relative humidity 50 % . Humidity calculated @ temperature 25°C
	Temperature		
	20 °C to 25 °C	1.2 °C	
	>25 °C to 30 °C	1.3 °C	
	Relative humidity		
	35 % to 50 %	3.1 %	
	>50 % to 65 %	4.1 %	
	Analog thermo - hygrometer / Thermo-hygrograph		In – House Method : CP-TEM-08A based on NPL : A guide to the measurement of humidity Temperature calculated @ relative humidity 50 % Humidity calculated @ temperature 25°C
	Temperature		
	20 °C to 25 °C	1.4 °C	
	>25 °C to 30 °C	1.5 °C	
	Relative humidity		
	35 % to 50 %	3.2 %	
>50 % to 65 %	4.2 %		
* express as an uncertainty (±), providing a level of confidence of approximately 95%			

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☐ Permanent ☒ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Range/Item	Calibration and Measurement Capability*	Method/Technique
1. Balance	Electronic balance and Mechanical balance up to 20 g > 20 g to 40 g > 40 g to 60 g > 60 g to 100 g > 100 g to 200 g > 200 g to 300 g > 300 g to 400 g > 400 g to 500 g > 500 g to 600 g > 600 g to 700 g > 700 g to 800 g > 800 g to 900 g > 900 g to 1 kg > 1 kg to 10 kg > 10 kg to 20 kg > 20 kg to 400 kg > 400 kg to 600 kg > 600 kg to 800 kg > 800 kg to 900 kg > 900 kg to 1 000 kg	0.13 mg 0.25 mg 0.26 mg 0.38 mg 0.65 mg 1.5 mg 1.6 mg 1.9 mg 2.1 mg 2.3 mg 2.5 mg 2.7 mg 3.0 mg 0.13 g 0.53 g 82 g 83 g 84 g 85 g 86 g	In - house method : CP- BAL-01A based on UKAS LAB 14 : 2006
* express as an uncertainty (\pm), providing a level of confidence of approximately 95%			

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☐ Permanent ☒ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Range/Item	Calibration and Measurement Capability*	Method/Technique
2. Mechanical	Pressure measuring instrument (Analog and digital) Pneumatic type 0 kPa to 200 kPa	0.19 kPa	In – house method : CP- PRE-01A based on DKD- R6-1,2003
	Water type 0 MPa to 70 MPa	58 kPa	In – house method : CP- PRE-02A based on DKD- R6-1,2003
	Vacuum measuring instrument (Analog and digital) -95 kPa to 0 kPa	0.33 kPa	In – house method : CP- PRE-03A based on DKD-R6-1,2003
3. Electrical	Temperature Indicator Thermocouple Type J -200 °C to 700 °C	0.40 °C	In - house Method : CP- TEM-02AS by direct measurement with documenting process calibrator based on EA10/11
	> 700 °C to 1 200 °C	0.39 °C	
	Type T 0 °C to 400 °C	0.33 °C	
	Type K -200 °C to 300 °C	0.40 °C	
	> 300 °C to 700 °C	0.37 °C	
	> 700 °C to 1 370 °C	0.38 °C	
	Type R 0 °C to 900 °C	1.2 °C	
	> 900 °C to 1 750 °C	0.70 °C	
* express as an uncertainty (±), providing a level of confidence of approximately 95%			

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☐ Permanent ☒ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Range/Item	Calibration and Measurement Capability*	Method/Technique
3. Electrical (cont.)	Temperature Indicator Resistance temperature detector Pt 100 Ω (385) 2, 3, 4 wire -200 °C to 800 °C	0.18 °C	In - house method : CP-TEM-22AS based on EA10/11 by direct measurement with documenting process calibrator
4. Temperature	Temperature sensor Thermocouple Type J -20 °C to 100 °C >100 °C to 200 °C >200 °C to 400 °C >400 °C to 600 °C Type K -20 °C to 100 °C >100 °C to 200 °C >200 °C to 300 °C >300 °C to 400 °C >400 °C to 500 °C >400 °C to 600 °C Type T -20 °C to 150 °C >150 °C to 250 °C	0.55 °C 0.89 °C 1.9 °C 2.9 °C 0.73 °C 0.99 °C 1.4 °C 1.7 °C 2.1 °C 2.4 °C 1.2 °C 1.7 °C	In - house method : CP-TEM-04AS by comparison with thermometer standard
* express as an uncertainty (\pm), providing a level of confidence of approximately 95%			

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☐ Permanent ☒ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Range/Item	Calibration and Measurement Capability*	Method/Technique
4. Temperature (cont.)	Temperature sensor		
	Resistance temperature detector		In - house method : CP-TEM-16AS
	(Pt 100 Ω) 2, 3, 4 wire		by comparison with thermometer standard
	-20 °C to 200 °C	0.25 °C	
	>200 °C to 300 °C	0.88 °C	
	>300 °C to 400 °C	1.3 °C	
	>400 °C to 500 °C	1.6 °C	
	>500 °C to 600 °C	1.9 °C	
	Temperature indicator with sensor		
	Thermocouple Type E, J, K, N, T		In - house method : CP-TEM-14AS by comparison with thermometer standard
	-20 °C to 100 °C	0.44 °C	
	>100 °C to 200 °C	0.76 °C	
	>200 °C to 300 °C	1.4 °C	
	>300 °C to 400 °C	1.9 °C	
	Thermocouple Type E, J, K, N		
	>400 °C to 500 °C	2.4 °C	
	>500 °C to 600 °C	2.8 °C	
	Resistance temperature detector		In - house Method : CP-TEM-12AS by comparison with thermometer standard
	(Pt 100 Ω) 2, 3, 4 wire		
	-20 °C to 200 °C	0.14 °C	
	>200 °C to 300 °C	0.38 °C	
	>300 °C to 400 °C	0.57 °C	
	>400 °C to 500 °C	0.86 °C	
	>500 °C to 600 °C	0.89 °C	
* express as an uncertainty (\pm), providing a level of confidence of approximately 95%			

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☐ Permanent ☒ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Range/Item	Calibration and Measurement Capability*	Method/Technique
4. Temperature (cont.)	Dial Thermometer		In - house method : CP-TEM-06AS by comparison with thermometer standard
	-20 °C to 200 °C	0.60 °C	
	>200 °C to 600 °C	1.1 °C	
	Liquid in glass thermometer		In - house method : CP-TEM-11AS by comparison with thermometer standard
	Total immersion		
	-20 °C to 100 °C	0.60 °C	
	Partial immersion		
	-20 °C to 100 °C	0.61 °C	
	Water bath		In - house method : CP-TEM-19AS based on ASTM E715-80 : (Reapproved 2001) by comparison with data acquisition
	20 °C to 80 °C	0.13 °C	
	Autoclave		In - house method : CP-TEM-20AS based on BS 2646 : 1993 Part 5 by comparison with data acquisition
	110 °C to 125 °C	0.60 °C	
	Hot air oven		In - house method : CP-TEM-21AS based on TLAS G-20 (Guidelines for calibration and check of temperature controlled enclosures)
	37 °C to 50 °C	0.40 °C	
	>50 °C to 100 °C	0.41 °C	
	>100 °C to 250 °C	0.52 °C	
* express as an uncertainty (\pm), providing a level of confidence of approximately 95%			

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☐ Permanent ☒ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Rage/Item	Calibration and Measurement Capability*	Method/Technique
4. Temperature (cont.)	Freezer -40 °C to 0 °C	0.52 °C	In - house method : CP-TEM-21AS based on TLAS G-20 (Guidelines for calibration and check of temperature controlled enclosures)
	Low Temp. Incubator 0 °C to 30 °C	0.33 °C	
	Incubator 30 °C to 100 °C	0.31 °C	
	Refrigerator 0 °C to 10 °C	0.33 °C	
* express as an uncertainty (±), providing a level of confidence of approximately 95%			

Scope of Accreditation for Calibration

Certificate No. 12132/C0217

Accreditation No. : Calibration 0217

Laboratory Status ☐ Permanent ☒ Site ☐ Temporary ☐ Mobile

Field of Calibration	Parameter/Range/Item	Calibration and Measurement Capability*	Method/Technique
4. Temperature (cont.)	Furnace 300 °C to 900 °C	3.6 °C	In - house method : CP-TEM-18AS based on BS 4309 : 1968 by comparison with Data acquisition
* express as an uncertainty (\pm), providing a level of confidence of approximately 95%			

Issue Date : 23 August B.E. 2555 (2012)

Signature :

(Urit Srinongkote)

Secretary – General

Thai Industrial Standards Institute